

Clenbuterol (CLE) ELISA Kit

Technical Manual

(ELISA)



Scan for more info.

Shenzhen Finder Biotech Co.,Ltd.
Web: www.szfinder.com
Tel: +86 0755 23499025 Email: techsupport@szfinder.com
Add: Building B12,Life Science Industrial Park, KuiyongSubdistrict,
Dapeng New Area, Shenzhen,China

1 Principle and Application

This kit adopts the method of indirect competitive enzyme-linked immunoassay (ELISA) to detect Clenbuterol (CLE) in the sample such as urine, tissue and feed. The kit is composed of Microtiter Plate coated with coupled antigens, HRP conjugate, antibodies, standards and other supporting reagents. During the detection, with adding standards or samples, the CLE in the samples will compete with the coupled antigens to combine with anti-CLE antibodies. After adding HRP conjugate, take coloration with TMB substrates. Absorbance value of the samples is a negative correlation with CLE content. Lastly, by comparing the obtained absorbance values with the standard curve, we can calculate the content of CLE content in the sample.

2 Technique Data

2.1 Kit Sensitivity: 0.05ppb (ng/mL)

2.2 Reactive Mode: 25°C, 30min ~ 15min

2.3 Detection Limits:

| Sample | Detection Limits |
|------------|------------------|
| Urine | 0.1ppb |
| Tissue | 0.2ppb |
| Feed | 1ppb |
| Pork liver | 0.5ppb |

2.4 Cross-reaction Rate:

| | |
|-------------|------|
| Clenbuterol | 100% |
| Terbutalin | <1% |
| Mabuterol | <1% |
| Brombuterol | <1% |
| Salbutamol | <1% |
| Ractopamine | <1% |

2.5 Sample Recovery Rate:

| Sample | Recovery rate |
|--------------------------|---------------|
| Urine | 95±10% |
| Tissue, feed, pork liver | 85±15% |

3 Composition of the Kit

| Reagent | Specification |
|---|------------------|
| Microtiter Plate | 8wells× 12strips |
| Standard: 0ppb, 0.05ppb, 0.15ppb, 0.45ppb, 1.35ppb, 4.05ppb (black cap) | 1.0mL each |
| High standard (red cap) | 100ppb |
| Antibody solution (blue cap) | 1×5.5mL |
| HRP conjugate (red cap) | 1×5.5mL |
| Substrate Reagent A (white cap) | 1×6mL |
| Substrate Reagent B (black cap) | 1×6mL |
| Stop Solution (yellow cap) | 1×6mL |
| Concentrated Wash Buffer (20×)(white cap) | 1×40mL |

| | |
|---|--------|
| Concentrated Reconstitution Buffer (10×) (yellow cap) | 1×50mL |
| Instruction | 1 |
| Adhesive Membrane | 1 |
| Sealed bag | 1 |

4 Materials Required but Not Supplied

4.1 Equipment: microplate reader, printer, grinder (for homogenizing solid samples), nitrogen evaporator, vortex mixer (for shake and mix), centrifuge, graduated transfer pipette, and balance with a division value of 0.01 g, constant temperature device;

4.2 Micropipette: single-channel (20-200μL, 100-1000μL), and multi-channel 300μL;

4.3 Reagents: Sodium Hydroxide, Ethyl Acetate, Concentrated Hydrochloric Acid, methanol, Anhydrous Sodium Carbonate.

5 Experimental preparation

Restore all reagents and samples to room temperature (adjust to around 25°C) for more than 30 min before use. This is a crucial step to ensure there is no precipitation in the reagents.

5.1 Notice Before Sample Processing:

Please note that the labware must be clean. Use disposable pipette tips to avoid contamination of interference results.

5.2 Solution preparation:

Solution 1: Reconstitution Buffer

Dilute the Concentrated Reconstitution Buffer (10×) 10 times (Concentrated Reconstitution Buffer (10×) /Deionized water= 1:9). It can be stored for one month at 4°C.

Solution 2: Working Wash Buffer

Dilute the concentrated wash buffer (20×) by a factor of 20 (Concentrated wash buffer/Deionized water= 1: 19).

Solution 3: Sample Extraction Solution

70% Methanol solution, (Methanol/Deionized water= 7: 3)

Solution 4: 10% Sodium Carbonate Solution

Dissolve 50 g of anhydrous sodium carbonate in deionized water and dilute to a total volume of 450 mL.

Solution 5: 0.1M Hydrochloric Acid Solution

Add 860 µL of concentrated hydrochloric acid to deionized water and dilute to a total volume of 100 mL.

Solution 6: 2M Sodium Hydroxide Solution

Dissolve 8 g of sodium hydroxide in deionized water and dilute to a total volume of 100 mL.

5.3 Sample pretreatment steps:

5.3.1 Urine treatment.

1) Take 1mL of clear urine sample (if the urine is turbid, filter it or centrifuge at 4000r/min for 5 minutes at room temperature to obtain a clear sample) in a 10mL centrifuge tube. Add 1mL of reconstitution buffer (solution 1), mix and shake for 30 seconds, then take 50µL for analysis. Samples not in use should be stored frozen.

Dilution times of the sample:2 Detection limits: 0.1ppb

5.3.2 Tissue treatment.

1) Weigh 1g ± 0.05g of homogenized tissue sample, add 3mL of reconstitution buffer (Solution 1), shake thoroughly for 2 minutes, and centrifuge at 4000r/min at room temperature for 10 minutes. (If the tissue sample has a high fat content, place it in an 85°C water bath for 10 minutes after shaking, then centrifuge.) Take 50µL of the supernatant for analysis.

Dilution times of the sample:4 Detection limits:0.2ppb

5.3.3 Animal feed treatment.

1) Weigh 1g ± 0.05g of ground feed sample into a 15mL centrifuge tube, add 4mL of sample extraction solution (solution 3), shake thoroughly for 2 minutes, and centrifuge at 4000r/min at room temperature for 10 minutes.

2) Transfer 0.1mL of the supernatant to a 1.5mL centrifuge tube, add 0.4mL of reconstitution buffer (solution 1), shake thoroughly for 30 seconds, and take 50µL for analysis.

Dilution times of the sample:20 Detection limits: 1ppb

5.3.4 Pork liver treatment.

1) Weigh 1g ± 0.05g of homogenized pig liver sample into a 15mL centrifuge tube, add 1mL of 10% sodium carbonate solution (solution 4), shake thoroughly for 2 minutes, add 4mL of ethyl acetate, shake thoroughly for 5 minutes, and centrifuge at 4000r/min at room temperature for 10 minutes

2) Transfer 2 mL of the supernatant to a 15 mL centrifuge tube, add 1 mL of 0.1M hydrochloric acid solution (Solution 5), vortex for 2 minutes to mix thoroughly, and then centrifuge at room temperature at 4000 rpm or higher for 5 minutes.

3) Transfer 0.1mL of the lower clear layer to a 1.5mL centrifuge tube, add 0.4mL of reconstitution buffer (solution 1), adjust the pH to around 7 using 2M sodium hydroxide solution (solution 6), and take 50µL for analysis.

Dilution times of the sample:10 Detection limits: 0.5ppb

6 ELISA procedure I

Place all reagents and samples to room temperature (adjust to around 25°C) for 30min. Gently shake the reagent bottles before use.

Take out the frame of the microplate along with the required number of wells. Then place the unused microplate wells into the sealed bag with the desiccant

provided. Store the remaining kit in the refrigerator at 2-8°C.

Step 1: Number: Number the wells in sequence corresponding to the samples and standard, make 2-well parallel trials for each sample and standard, and record their locations.

Step 2: Incubation: Add 50µL of standard or sample into each numbered well, then add 50µL of HRP conjugate per well. Next, add 50µL of antibody solution per well. Finally, cover the Microtiter Plate with the adhesive membrane, shake gently by hand (or use a microplate shaker) for 5s and incubate for 30 min at 25°C.

Step 3: Washing: Uncover the adhesive membrane carefully, remove the liquid, pipette 350µL of Working Wash Buffer (Solution 2) to every well, let stand for 30 seconds then drain, repeat 5 times. Invert the plate and tap it against a thick absorbent paper (or lint-free cloth), with a soft towel placed underneath. (Bubbles that are not removed after patting dry can be punctured with a clean pipette tip).

Step 4: Color: Add 50µL of Substrate Reagent A to each well, then add 50µL of Substrate Reagent B per well. Shake gently by hand (or use a microplate shaker) for 5s, and allow to react for 15min at 25°C in the dark. (The reaction can be extended appropriately if the blue colour is too pale.)

Step 5: Stop the reaction: Pipette 50µL of Stop Solution to each well, shake gently by hand (or use a microplate shaker) for 5s.

The reaction would be stopped.

Step 6: Calculate: Determine the Optical Density (OD value; absorbance value) at 450nm (Reference wavelength 630nm) with a microplate reader. Finish this step within 10min after stop the reaction.

7 Interpretation of result |

7.1 Calculate the percentage of absorbance value

Percentage of absorbance value(%)= $\frac{A}{A0} \times 100\%$

A—the average OD value of the sample or standard;

A0—the average OD value of the 0ppb standard.

It is used to calculate the percentage absorbance of a standard or sample.

7.2 Draw the standard curve and calculate

Take absorbance percentage (A/A0) of standards as Y-axis and the corresponding log of standards concentration (ppb) as X-axis.

Draw the standard semi-log curves with X-axis and Y-axis.

Take absorbance percentage of samples substitute into standard curve, then can get the corresponding concentration from standard curve. **Last, the resulting concentration values multiplied by the corresponding dilution times is the actual concentration of CLE of samples.**

If professional analysis software of the kit is used for calculation, it is more convenient for accurate and rapid analysis of a large number of samples.

8 Attention |

8.1 Before test, the reagents and samples should be balanced to room temperature (25°C). If below 25°C, it will lead to all the standard OD value on the low side.

8.2 In the washing process, dry wells may result in non-linear standard curves and undesirable reproducibility. Therefore, proceed to the next step immediately after washing.

8.3 Please mix the contents within the wells uniformly and wash the plate thoroughly. The reproducibility is

largely determined by consistency of washing step.

8.4 During the incubation, cover microplates with adhesive membrane to avoid light.

8.5 Do not use kits that are overdue. Do not mix reagents with those from other lots.

8.6 Substrate Reagent A/B is colorless. If not, please discard.

8.7 If absorbance value of 0ppb is below 0.5 (A450nm< 0.5), it means that the reagent may be metamorphic.

8.8 Stop solution is corrosives, please avoid contact with skin.

8.9 As the OD values of the standard curve may vary according to the conditions of actual assay performance (e.g. operator, pipetting technique, washing technique or temperature effects), the operator should establish a standard curve for each test.

8.10 For the mentioned sample, fast and efficient extraction methods are included in the kit description. Please consult technical support for the applicability if other sample need to be tested.

8.11 The kit is used for rapid screening of actual samples. If the test result is positive, the instrument method such as HPLC, LC/MS can be used for quantitative confirmation.

9 Storage conditions |

The kit shall be stored at 2-8 °C. Avoid freezing.

Shelf Life: 12 months. The date of manufacture is presented in the label of the box.