

# Bovine Hydatidosis Antibody (HYD Ab) ELISA Kit

## Technical Manual

(ELISA)



Scan for more info.

Shenzhen Finder Biotech Co.,Ltd.  
Web: [www.szfinder.com](http://www.szfinder.com)  
Tel: +86 0755 23499025 Email: [techsupport@szfinder.com](mailto:techsupport@szfinder.com)  
Add: Building B12,Life Science Industrial Park, KuiyongSubdistrict,  
Dapeng New Area, Shenzhen,China

## Product Information

### Intended Use

Hydatid disease, also known as echinococcosis, is a chronic parasitic infection resulting from the larvae of *Echinococcus granulosus* and *Echinococcus multilocularis*. This assay is designed for the detection of hydatid antibodies in bovine serum. It can be used for the evaluation of the immunogenic efficacy of hydatid vaccines or as an adjunctive diagnostic tool for the disease.

This assay is designed to detect hydatid antibodies in the bovine serum or plasma. It can be utilized for the evaluation of the immunological efficacy of hydatid vaccines or as an adjunctive diagnostic tool for the disease.

### Principle and Application

This kit comprises a coated Microtiter Plate with hydatid antigens, HRP conjugate, and other

accompanying reagents. It employs the principle of enzyme-linked immunosorbent assay (ELISA) to detect hydatid antibodies in the bovine serum or plasma. During the experiment, control serum and test sample are added to the plate. After incubation, if the sample contains hydatid antibodies, they will bind to the antigens coated on the microtiter plate. Following washing steps to remove unbound components, the HRP conjugate is added, which specifically binds to the antigen-antibody complexes on the plate. After washing again to remove unbound HRP conjugate, substrate reagents are added to the wells and react with the enzyme-labeled complexes, resulting in a blue color. The intensity of the color is directly proportional to the amount of specific antibody present in the sample. The reaction is then terminated by adding a stop solution, turning the solution yellow. The absorbance of each well is measured at a wavelength of 450nm using a microtiter plate reader (microplate reader) to determine the presence of antibodies against hydatid in the sample.

### Composition of the Kit

Reagent	Specification		
Microtiter Plate	96wells	96wells×2	96wells×5
Sample Diluent (yellow cap)	1×50mL	1×50mL	1×200mL
HRP conjugate (red cap)	1×11mL	2×11mL	2×26mL
Concentrated Wash Buffer (20×) (white cap)	1×40mL	1×40mL	1×200mL
Substrate Reagent A (white cap)	1×6mL	1×11mL	1×26mL
Substrate Reagent B (black cap)	1×6mL	1×11mL (brown cap)	1×26mL
Stop Solution (yellow cap)	1×6mL	1×11mL	1×26mL
Positive Control (red cap)	1×1.0mL	1×1.5mL	1×2.0mL
Negative Control (green cap)	1×1.0mL	1×1.5mL	1×2.0mL
Adhesive Membrane	1	2	5
Sealed bag	1	1	2
Dilution plate	1	2	5
Instruction	1	1	1

### Storage conditions

The kit shall be stored at 2-8 °C. Avoid moisture.

Shelf life: 12 months. Please use within 2 months after opening. The date of manufacture is presented in the label of the box.

### Test Apparatus Required but Not Supplied

microplate reader, adjustable micropipette, constant temperature device (37°C), centrifuge.

## Experimental preparation

**Restore all reagents and samples to room temperature (adjust to around 25°C) for more than 30 min before use. This is a crucial step to ensure there is no precipitation in the reagents.**

**1.Sample Preparation:**The serum/plasma should be clear, without hemolysis or contamination. Samples can be stored at 2-8°C for up to 1 week, and for long-term storage, they should be kept at -20°C.

**2.Sample dilution:** Dilute the prepared sample 10-fold using the sample diluent (e.g., add 90μL of sample diluent to the dilution plate, followed by the addition of 10μL of the sample, and mix well). What we get is the diluted sample. Negative and positive controls do not require dilution.

**3.Solution preparation:** Dilute the concentrated wash buffer (20×) by a factor of 20 (Concentrated wash buffer/Deionized water= 1: 19). What we get is the working wash buffer.

Please note that the labware must be clean. Use disposable pipette tips to avoid contamination of interference results.

## ELISA procedure

Place all reagents and samples to room temperature

(adjust to around 25°C) for 30min. Gently shake the reagent bottles before use.

Take out the frame of the microplate along with the required number of wells. Then place the unused microplate wells into the sealed bag with the desiccant provided. Store the remaining kit in the refrigerator at 2-8°C.

1.Put the required number of the wells on the plate and set up 2 wells each for negative/positive control.

2.Add 100μL of **negative control** to each negative control well. Then add 100μL of **positive control** to each positive control well. For each sample well, first add 90μL of **sample diluent**, then add 10μL of the **diluted sample**.

**Note: The sample dilution ratio is equivalent to 1:100.**

3.Shake the plate gently by hand (or use a microplate shaker) for 5s, cover with adhesive membrane and incubate at 37°C (water bath recommended) in the dark for 30 minutes.

4.Discard the liquid from the wells. Add 350μL of **working wash buffer** to each well, let stand for 30 seconds, then discard. Repeat the washing process 5 times. Invert the plate and tap it against a thick absorbent paper (or lint-free cloth), with a soft towel placed underneath. (Bubbles that are not removed after tapping dry can be punctured with a clean pipette tip).

5.Add 100μL of **HRP conjugate** to each well, cover with adhesive membrane, and incubate at 37°C in the dark for 30 minutes.

6.Washing. Same as step 4.

7.First, add 50μL of **substrate reagent A** to each well, followed by 50μL of **substrate reagent B**. shake gently by hand (or use a microplate shaker) for 5s, cover with adhesive membrane, and incubate at 37°C in the dark for 10 minutes.

8.Add 50μL of **stop solution** to each well and shake gently by hand (or use a microplate shaker) for 5s. Read absorbance (**A value**) at 450nm with microplate reader (with 630nm as a reference wavelength). Finish this step within 10min.

#### Reference Value

Under normal experimental conditions, the A value of the negative control should be  $\leq 0.15$ , and the A value of the positive control should be  $\geq 0.6$ .

#### Interpretation of Test Results

1. If  $A_s \geq 0.5$ , it is considered positive; if  $A_s < 0.5$ , it is considered negative.

$A_s$ —the A value of the sample

2 When the result is positive, it indicates the presence of antibodies against hydatid in bovine. Further analysis should be conducted in conjunction with clinical and other methods.

#### Limitations of the Test Method

This test result is for screening purposes only and is not sufficient for confirmation.

#### I Attention I

1. During the experiment, gloves and lab coats should be worn. Strict and comprehensive disinfection and isolation protocols should be followed. All experimental waste should be treated as infectious material.

2. The stop solution is corrosive. Avoid contact with skin and clothing. If accidentally contacted, rinse immediately with a large amount of tap water.

3. When taking the microtiter plate out of a refrigerated environment, it should be brought to room temperature before opening the bag. Unused microplate wells should be stored in the sealed bag with a desiccant.

4. During washing, each well should be filled completely with liquid to prevent any residual enzyme on the well's rim

from remaining unwashed.

5. The samples used for testing should be kept fresh.

6. The determination of test results must be based on the readings from the microplate reader.

7. Components from different lot numbers must not be mixed.